



25 February 2025

SUPPLEMENTAL BID BULLETIN
ADDENDUM NO. 2025-08-01

Subject: Supply and Delivery of Various Meteorological Instruments
(Reference: PR 2025-01-0055 || IB 2025-08)

This Bid Bulletin is being issued to all prospective bidders to clarify, amend and/or modify certain provisions in the Bidding Documents and to answer written queries and clarificatory questions during the pre-bid conference by prospective bidders, to wit:

Queries from: **Sea and Land Technologies (Philippines) Inc.**

Queries	Answers
<p>Single Largest Completed Contract</p> <p>Supply and Delivery of Various Meteorological Instruments completed within three (3) years prior to the deadline for submission of bids.</p> <p><u>Request for Amendment</u></p> <p>Supply and Delivery of Survey and Scientific Equipment (ex. Hydrological) completed within three (3) years prior to the deadline for submission of bids.</p>	<p>The current requirement for the Supply and Delivery of Various Meteorological Instruments mandates that the bidder must have successfully completed similar projects within three (3) years prior to the deadline for the submission of bids. This is a standard practice to ensure the bidder has recent, relevant experience and can meet the precise specifications for meteorological instruments, which are critical for accurate weather monitoring at Synoptic and Agromet Stations.</p> <p>The request for amendment of the SLCC requirement cannot be granted since the requirement as stated allows the agency to determine and evaluate whether the bidder has the expertise and experience in providing meteorological instruments which will be helpful and effective in weather observation.</p>

Queries from: **Metronix Inc.**

Queries	Answers
<p>Calibration Test</p> <p>Is the Calibration Test with documentary proof of calibration from PAGASA are included in the Delivery Schedule?</p>	<p>The Delivery Schedule pertains only to the supply and delivery of the equipment. The calibration test and certificate can be provided after the delivery, before the official acceptance of the equipment.</p>
<p>Assmann Aspiration Psychrometer</p> <p>In Assmann Aspiration Psychrometer referring to the accessories a “compact stand” is this the Standard Stand that included in the set?</p>	<p>Yes, the compact stand is the accessory included in the set.</p>

This shall form an integral part of the Bid Documents.

For guidance and information of all participating bidders.

Sgd.
ATTY. ERIC CHRISTOPHER AMADO J. AGSAOAY
Vice-Chairperson, PAGASA-BAC

“tracking the sky...helping the country”