



13 April 2026

SUPPLEMENTAL BID BULLETIN
ADDENDUM NO. 2026-29-01

Subject: Supply, Delivery, Installation, Configuration and Testing of Automatic Weather Stations with Agroment Sensors - Re Bid (Reference: PR 2026-01-0021 || IB 2026-29 R)

This Bid Bulletin is being issued to all prospective bidders to clarify, amend and/or modify certain provisions in the Bidding Documents and to answer written queries and clarificatory questions during the pre-bid conference by prospective bidders, to wit:

Queries from: **Maedan Enterprise Inc.**

Queries	Answers
<p>Clarification on the datalogger specification, particularly the reference to <i>LoggerNet</i>, as it appears to be associated with a specific brand, and whether equivalent data logging systems from other manufacturers may be considered acceptable</p>	<p>To ensure the continued homogenization of our dataloggers and sensors across the network, wherein the same brand and model are consistently used, please be informed that the specification is intentionally aligned with the existing system.</p> <p>This requirement is in accordance with the standardization of meteorological sensors and components (e.g., dataloggers) implemented by PAGASA to minimize instrument bias across its observation network and to ensure data consistency and accuracy.</p> <p>Such standardization is aligned with internationally recognized practices, including those recommended by the World Meteorological Organization (WMO).</p>

This shall form an integral part of the Bid Documents.

For guidance and information of all participating bidders.

sgd.
SHIRLEY J. DAVID, MoS
Chairperson, PAGASA-BAC

"tracking the sky...helping the country"