

Republic of the Philippines

DEPARTMENT OF SCIENCE AND TECHNOLOGY

Philippine Atmospheric, Geophysical and Astronomical Services

Administration (PAGASA)

06 September 2021

SUPPLEMENTAL / BID BULLETIN

ADDENDUM NO. 2021-038-01

Subject: Supply, Delivery and Testing of Minimum Thermometer (Alcohol-Filled) Mounted on Wood / ABC Php 1,152,800.00 / PR No. 2021-08-0546 / Ref. No. IB 2021-038


This Bid Bulletin is being issued to all prospective bidders to clarify, amend and/or modify certain provisions in the Bidding Documents and to answer the written clarificatory questions submitted by prospective bidders, to wit:

Query from **Clean World Trading**

Question(s)	PAGASA BAC Answer(s)
<p>Technical Specifications:</p> <ul style="list-style-type: none">- Alcohol Filled- Measuring range: -25 to +50°C- Scale Graduation: 0.5°- Overall length: 315 to 345mm- Capillary yellow back- Capillary in an outer glass sheath- Mounted on wood- Calibration: ISO 9001 compliant factory calibration <p>Clarification <i>It should be:</i></p> <ul style="list-style-type: none">- Alcohol Filled- Measuring range: -25 to +50°C- Scale Graduation: 0.5°- Overall length: 315 to 345mm- Capillary yellow back- Capillary in an outer glass sheath- Mounted on wood / Wooden base- And Stainless-steel plate for mounting (to be place in instrument shelter)- Calibration: ISO 9001 compliant factory calibration	<p>Yes. The two (2) pcs stainless-steel plate is needed for mounting and it will be placed in the instrument shelter.</p>

This shall form an integral part of the Bid Documents.

For guidance and information of all participating bidders.


LANDRICO U. LALIDA JR., Ph.D.
Chairperson
PAGASA-BAC for Goods and Consulting Services